

# UK Declaration of Conformity

The manufacturer:

**Kistler Instrumente AG, Eulachstrasse 22, CH-8408 Winterthur, Switzerland**

hereby declares that the product

Name	<b>KiBox2</b>	
Type	<b>2895A..</b>	
Modules	<b>4667A 5270A / 5270B 5075A.</b>	<b>PRAQ-Modul VAQ-Modul PEAQ-Modul</b>
Options	<b>all</b>	

complies with the following provisions of directives:

<b>UK SI 2017 No. 1206</b>	<b>(RED)</b>
<b>UK SI 2012 No. 3032</b>	<b>(ROHS)</b>

The following harmonised standards were applied:

EMC Emission

<b>EN 301 489-1</b>	<b>V2.2.3</b>
<b>EN 301 489-17</b>	<b>V3.2.4</b>
<b>EN 301 489-19</b>	<b>V2.1.1</b>
<b>EN 61326-1:2013</b>	<b>(Class A)</b>

Effective use of radio frequencies

<b>EN 300 328</b>	<b>V2.2.2</b>
-------------------	---------------

Low Voltage

<b>EN 61010-1:2010</b>
------------------------

**Kistler Instrumente AG, Winterthur**

Winterthur, 09. Nov 2022



**C. Gerber**  
Product Compliance Coordinator



**R. Odermatt**  
Head of Product Care Electronics