



Schweizerische Eidgenossenschaft
Confédération suisse
Confederazione Svizzera
Confederaziun svizra

Federal Department of Economic Affairs,
Education and Research EAER
State Secretariat for Economic Affairs SECO
Swiss Accreditation Service SAS

Swiss Confederation

Based on the Accreditation and Designation Ordinance dated 17 June 1996 and on the advice of the Federal Accreditation Commission, the Swiss Accreditation Service (SAS) grants to

Kistler Instrumente AG
SCS Calibration Laboratory
Eulachstrasse 22
8408 Winterthur



Period of accreditation:
28.08.2021 until 27.08.2026
(1st accreditation: 05.10.1994)

the accreditation as

Calibration laboratory for pressure, force and electrical quantities

International standard: ISO/IEC 17025:2017

Swiss standard: SN EN ISO/IEC 17025:2018

3003 Berne, 24.06.2021
Swiss Accreditation Service SAS

Head of SAS
Konrad Flück

SAS is a signatory of the multilateral agreements of the European co-operation for Accreditation (EA) for the fields of testing, calibration, inspection and certification of management systems, certification of personnel and certification of products, processes and services, of the International Accreditation Forum (IAF) for the fields of certification of management systems and certification of products, processes and services and of the International Laboratory Accreditation Cooperation (ILAC) for the fields of testing, calibration and inspection.



SCS Directory

Accreditation number: SCS 0049

International standard: ISO/IEC 17025:2017
Swiss standard: SN EN ISO/IEC 17025:2018

Kistler Instrumente AG
SCS Calibration Laboratory
Eulachstrasse 22
Postfach
8408 Winterthur

Head: Dr Georg Schading
Responsible for MS: Thomas Treffler
Telephone: +41 52 224 11 11
E-Mail: accreditation@kistler.com
Internet: www.kistler.com
Initial accreditation: 05.10.1994
Current accreditation: 28.08.2021 to 27.08.2026
Scope of accreditation see: www.sas.admin.ch
(Accredited bodies)

Scope of accreditation as of 28.08.2021

Calibration laboratory for pressure, force and electrical quantities

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Overpressure in fluids Calibration of piezoelectric pressure sensors	1 ... < 10 bar	stepwise	0,06 %	above 1000 bar with pressure multiplier
	10 ... < 100 bar	change of	0,03 %	
	100 ... 1000 bar	pressure load	0,01 %	
	1000 ... 8000 bar		0,05 %	
Overpressure in fluids Calibration of piezoresistive pressure sensors	0 ... < 5 bar	stepwise	0,1 %	above 1000 bar with pressure multiplier
	5 ... < 50 bar	change of	0,03 %	
	50 ... 1000 bar	pressure load	0,01 %	
	1000 ... 5000 bar		0,05 %	



SCS Directory

Accreditation number: SCS 0049

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Force Calibration of piezoelectric force sensors	0,05 ... < 2 kN	stepwise / continuous change of force load	0,2 %, but not less than 0,4 N	50 kN K-BNME
	2 ... 50kN		0,15 %	
	1 ... 100 kN		0,2 %	100 kN K-BNME
	1 ... < 50 kN		0,2 %	300 kN K-BNME
	50 ... 300 kN		0,15 %	
	10 ... < 50 kN		0,2 %	
Electrical charge Generation and calibration	50 ... 500 kN	0,15 %	500 kN K-BNME	
	1 ... < 20 pC	0,007 pC		
	20 ... < 50 pC	80 ppm + 0,006 pC		
	50 ... < 200 pC	170 ppm		
	200 ... < 48000 pC	150 ppm		
Voltage (DC)	48 ... 3100 nC	190 ppm		
	0 ... < 0,12 V	6,8 ppm + 2,7 μ V		
	0,12 ... < 1,2 V	14,2 ppm + 4,3 μ V		
	1,2 ... < 12 V	17,5 ppm + 2,7 μ V		
Voltage (AC)	12 ... < 100 V	14,2 ppm + 387 μ V		
	0 ... < 0,12 V	1 Hz ... 1 kHz	251 μ V	
	0,12 ... < 1,2 V	1 Hz ... 1 kHz	15 ppm + 264 μ V	
	1,2 ... < 12 V	1 Hz ... 1 kHz	51 ppm + 516 μ V	
	12 ... < 30 V	1 Hz ... 1 kHz	150 ppm + 5,1 mV	
	0 ... < 0,33 Vpp	1 Hz ... 1 kHz	708 μ Vpp	
	0,33 ... < 3,3 Vpp	1 Hz ... 1 kHz	53 ppm + 723 μ Vpp	
	3,3 ... < 33 Vpp	1 Hz ... 1 kHz	130 ppm+1,2 mVpp	
33 ... 85 Vpp	1 Hz ... 1 kHz	188 ppm+ 14 mVpp		



SCS Directory

Accreditation number: SCS 0049

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Current (DC)	0 ... < 0,37 mA		4,6 ppm + 34 nA	
	0,37 ... < 1,4 mA		23,6 ppm + 27 nA	
	1,4 ... < 4,5 mA		28 ppm + 50 nA	
	4,5 ... < 144 mA		35,4 ppm + 14,5 nA	
	144 ... 1000 mA		32,7 ppm + 3,2 μ A	
Resistance (DC)	0,01 ... < 12 Ω		19,7 ppm + 122 $\mu\Omega$	
	12 ... < 120 Ω		19,4 ppm + 1,2 m Ω	
	0,12 ... < 1,2 k Ω		15,3 ppm + 1,1 m Ω	
	1,2 ... < 12 k Ω		15,3 ppm + 11 m Ω	
	12 ... < 120 k Ω		16 ppm + 100 m Ω	
	0,12 ... < 1,2 M Ω		20 ppm + 4,1 Ω	
	1,2 ... < 12 M Ω		75 ppm + 102 Ω	
	12 ... 120 M Ω		0,1 % + 1,8 k Ω	
Capacitance	1 ... < 1000 pF	1 kHz	29 ppm	
	1 ... < 100 nF	1 kHz	34 ppm	
	100 ... < 1000 nF	1 kHz	113 ppm	
	1 ... < 10 pF	50 Hz ... 20 kHz	85 ppm	
	10 ... < 100 pF	50 Hz ... 20 kHz	41 ppm	
	100 ... < 1000 pF	50 Hz ... 20 kHz	34 ppm	
	1 ... < 10 nF	50 Hz ... 20 kHz	34 ppm	
	10 ... < 100 nF	50 Hz ... 20 kHz	123 ppm	
	100 ... 1000 nF	50 Hz ... 10 kHz	455 ppm	

* / * / * / * / *